


# Flow Measurement Test Kits



# Flow Measurement Test Kits

Model 300.4 & 300.3





# Flow Measurement Test Kit Model 300.4

Model 300.4 Test Kit is a portable unit for field monitoring and adjusting of AccuSetter (FlowSet) valves, Pitot Tubes and Venturis. The kit contains a  $\pm 1.75\%$  F.S. accurate, 6" diameter face, 370° arc. berylliumdiaphram style gauge with three-valve manifold for over-range protection. Unit is housed in a vinyl case and comes equipped with ten foot hoses and can read either temperature/pressure ports or pressure only access valves. Use Flowcharts F234, or F193 in conjunction with label located on valve.



# **Key features**

- > Large Face
  Easy reading
- > 1.75% FS Accuracy

> 3-Valve Manifold Protects meter during connection

# **Technical description**

# **Application:**

Reading Venturis

## **Functions:**

Pressure Reading

#### **Case Dimensions:**

14"L x 12 1/2"W x 6"D

## Weight:

11 1/4 lbs.

# **Description:**

6" Diameter Face 10' Hose length P/T and Schrader Adapters Dual Scales Feet and Inches W.C. ± 1 3/4% Special Calibration Beryllium Diaphragm

# Pressure class:

500 psi working pressure

# Maximum temperature:

200°

# Accuracy:

±1 3/4% special calibration

#### Material:

Body: Brass

# **Options**

Dial Range (Specify):

- □ 0 100" / 0 8.33' Std. FlowSet
- □ 0 300" / 0 25' Std. FlowSet
- □ 0 500" / 0 40' Std. FlowSet

# Model 300.3

Model 300.3 Test Kit is a differential pressure meter designed for permanent installation and consists of a 4 1/2" diaphragm type meter (=2% F.S.). Available to read directly in GPM by special order only. Uses customer supplied piping, valves, etc. for field attachment to the flow element. The flow element must be specified at the time of order.



# **Key features**

> Direct GPM Reading
Ease of use

> 4 1/2" Face
Good resolution

# **Technical description**

# **Application:**

Dedicated Meter for Flow Reading

# **Functions:**

Flow Measurement

# Weight:

4 1/2 lbs.

# **Description:**

4.5" or 6" Diameter Face No Hoses (for direct mounting) ± 2% Special Calibration Buna-N Diaphragm

# Pressure class:

500 psi working pressure

## Maximum temperature:

210°

# Accuracy:

±2% special calibration

# Material:

Body: Brass

# **Options**


# Model 300.4 Dial Range (Specify):

- □ 0 100" / 0 8.33' Std. FlowSet
- □ 0 300" / 0 25' Std. FlowSet
- □ 0 500" / 0 40' Std. FlowSet

# Model 300.3 Must Specify

(Device to be read and GPM)

# Installation



- 1. Hang the meter by its support cable so that the dial face is level.
- 2. Close valves (1) & (2). Open valve (5). Close valves (6) & (7).
- 3. Close valves on Red & Yellow hoses. Connect high pressure fitting (8) to upstream orifice tap (Red hose) and connect low pressure fitting (9) to down stream orifice tap (Yellow hose) using rubber hoses provided. The hose valves should be located close to the venturi tap.
- 4. Open the hose valves and any valves on the venturi impulse lines.
- 5. Open valves (6) & (7) and alternately crack valves (1) & (2) until all air has been expelled from the instrument & hoses. Close valves (1) & (2).
- 6. Open valves (6) & (7), slowly close by-pass valve (5) make sure the meter is not over-ranged. Read differential pressure.
- 7. When through with test, open valve (5), and close hose valves. Remove hose from venturi. DO NOT REMOVE HOSES FROM METER IF ANOTHER VENTURI IS TO BE MEASURED.
- 8. The filled and purged meter & hoses can be used for another test without purging.

# Installation

Normal operation requires no maintenence. In cases where dirt or scale builds up, flush the meter with clean water or a compatible solvent.

# **CAUTION!**

Do not leave liquid in the meter below freezing temperatures.

Care must be taken not to overpressure the meter.

Secure and support the hoses to prevent sagging and/or vibration.



